

---

---

**Tobacco — Determination of the content  
of total alkaloids as nicotine —  
Continuous-flow analysis method**

*Tabac — Détermination de la teneur en alcaloïdes totaux exprimés en  
nicotine — Méthode par analyse en flux continu*

**PDF disclaimer**

This PDF file may contain embedded typefaces. In accordance with Adobe's licensing policy, this file may be printed or viewed but shall not be edited unless the typefaces which are embedded are licensed to and installed on the computer performing the editing. In downloading this file, parties accept therein the responsibility of not infringing Adobe's licensing policy. The ISO Central Secretariat accepts no liability in this area.

Adobe is a trademark of Adobe Systems Incorporated.

Details of the software products used to create this PDF file can be found in the General Info relative to the file; the PDF-creation parameters were optimized for printing. Every care has been taken to ensure that the file is suitable for use by ISO member bodies. In the unlikely event that a problem relating to it is found, please inform the Central Secretariat at the address given below.

This document is a preview generated by EVS

© ISO 2003

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office  
Case postale 56 • CH-1211 Geneva 20  
Tel. + 41 22 749 01 11  
Fax + 41 22 749 09 47  
E-mail [copyright@iso.org](mailto:copyright@iso.org)  
Web [www.iso.org](http://www.iso.org)

Published in Switzerland

**Contents**

	Page
<b>1 Scope</b> .....	<b>1</b>
<b>2 Normative references</b> .....	<b>1</b>
<b>3 Principle</b> .....	<b>1</b>
<b>4 Safety precautions</b> .....	<b>1</b>
<b>5 Reagents</b> .....	<b>4</b>
<b>6 Apparatus</b> .....	<b>5</b>
<b>7 Procedure</b> .....	<b>5</b>
<b>8 Calculation</b> .....	<b>6</b>
<b>9 Repeatability and reproducibility</b> .....	<b>6</b>
<b>10 Test report</b> .....	<b>7</b>
<b>Annex A (informative) Information about alternative analytical methods</b> .....	<b>8</b>
<b>Annex B (informative) Example of a continuous-flow analyser</b> .....	<b>9</b>
<b>Annex C (informative) Preparation of a microbore five-turn mixing coil</b> .....	<b>10</b>
<b>Bibliography</b> .....	<b>11</b>

## Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

International Standards are drafted in accordance with the rules given in the ISO/IEC Directives, Part 2.

The main task of technical committees is to prepare International Standards. Draft International Standards adopted by the technical committees are circulated to the member bodies for voting. Publication as an International Standard requires approval by at least 75 % of the member bodies casting a vote.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights.

ISO 15152 was prepared by Technical Committee ISO/TC 126, *Tobacco and tobacco products*, Subcommittee SC 2, *Leaf tobacco*.

This document is a preview generated by EVS

## Introduction

A CORESTA<sup>1)</sup> Task Force studied the various widely used procedures for the determination of total alkaloids in tobacco in order to adopt one of them as the CORESTA Recommended Method. Two procedures were adopted as ISO 2881 and this International Standard. Studies carried out by the CORESTA Task Force between 1989 and 1993 have shown that the two methods may not produce identical results for some dark tobaccos or those containing significant levels of alkaloids other than nicotine. The studies have indicated that these differences may be due to the fact that the recoveries and detection sensitivities of the methods for alkaloids other than nicotine are different.

---

1) CORESTA: Cooperation Centre for Scientific Research Relative to Tobacco.

This document is a preview generated by EVS

# Tobacco — Determination of the content of total alkaloids as nicotine — Continuous-flow analysis method

## 1 Scope

This International Standard specifies a method for the determination of the content of total alkaloids as nicotine in tobacco by continuous-flow analysis.

This method is applicable to manufactured and unmanufactured tobacco.

## 2 Normative references

The following referenced documents are indispensable for the application of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 13276, *Tobacco and tobacco products — Determination of nicotine purity — Gravimetric method using tungstosilicic acid*

## 3 Principle

An aqueous extract (see below) of the tobacco is prepared and the total alkaloids content (as nicotine) of the extract is determined by reaction with sulfanilic acid and cyanogen chloride. Cyanogen chloride is generated *in situ* by the reaction of potassium cyanide and chloramine T (see Annex A). The developed colour is measured at 460 nm.

Collaborative studies have shown that this method gives equivalent results for water and 5 % acetic acid extracts. It is recommended that 5 % acetic acid extracts should be used if total alkaloids (as nicotine) and reducing substances (see ISO 15153) or reducing carbohydrate analyses (see ISO 15154) are to be carried out simultaneously.

## 4 Safety precautions

**WARNING — Potassium cyanide is poisonous and an irritant, thus all safety precautions shall be observed when handling this material.**

Solutions shall be prepared by a designated responsible person. Gloves and safety glasses shall always be used when making up solutions. Bottles of the made-up reagent shall always be carried in a suitable safety carrier. To prevent the escape of vapour into the laboratory, reagent pick-up tubes shall pass through a soda-lime trap into the reagent bottle (see Figure 1).

The cyanide neutralizing solutions A and B (5.5 and 5.6) shall be pumped as shown in the flow diagram (see Figure B.1) and mixed in a 2 litre Büchner flask with magnetic stirring (see Figure 2). All waste solutions containing cyanogen chloride shall be run into this flask where conversion to the "Prussian Blue" complex occurs. The contents of the Büchner flask shall be allowed to overflow into a storage flask, the contents of which shall be stored overnight in a fume cupboard and then disposed of as waste.

Suitable cyanide poisoning treatment kits are available from laboratory equipment suppliers and shall be located in the vicinity of the analyser to be used by a competent person.