

VEE KVALITEET. PERMANGANAATARVU MÄÄRAMINE

Water quality - Determination of permanganate index

This document is a preview generated by EVS

EESTI STANDARDI EESSÕNA**NATIONAL FOREWORD**

See Eesti standard EVS-EN ISO 8467:1999 sisaldab Euroopa standardi EN ISO 8467:1995 ingliskeelset teksti.	This Estonian standard EVS-EN ISO 8467:1999 consists of the English text of the European standard EN ISO 8467:1995.
Standard on jõustunud sellekohase teate avaldamisega EVS Teatajas.	This standard has been endorsed with a notification published in the official bulletin of the Estonian Centre for Standardisation.
Euroopa standardimisorganisatsioonid on teinud Euroopa standardi rahvuslikele liikmetele kättesaadavaks 27.03.1995.	Date of Availability of the European standard is 27.03.1995.
Standard on kättesaadav Eesti Standardikeskusest.	The standard is available from the Estonian Centre for Standardisation.

Tagasisidet standardi sisu kohta on võimalik edastada, kasutades EVS-i veebilehel asuvat tagasiside vormi või saates e-kirja meiliaadressile standardiosakond@evs.ee.

ICS 13.060.50

Standardite reprodutseerimise ja levitamise õigus kuulub Eesti Standardikeskusele

Andmete paljundamine, taastekitamine, kopeerimine, salvestamine elektroonsesse süsteemi või edastamine ükskõik millises vormis või millisel teel ilma Eesti Standardikeskuse kirjaliku loata on keelatud.

Kui Teil on küsimusi standardite autorikaitse kohta, võtke palun ühendust Eesti Standardikeskusega:
Koduleht www.evs.ee; telefon 605 5050; e-post info@evs.ee

The right to reproduce and distribute standards belongs to the Estonian Centre for Standardisation

No part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying, without a written permission from the Estonian Centre for Standardisation.

If you have any questions about copyright, please contact Estonian Centre for Standardisation:

Homepage www.evs.ee; phone +372 605 5050; e-mail info@evs.ee

ICS 13.060.40

Descriptors: Water quality, water analysis, permanganate index. testing.

English version

Water quality

**Determination of permanganate index
(ISO 8467:1993)**

Qualité de l'eau; détermination de l'indice
de permanganate (ISO 8467:1993)

Wasserbeschaffenheit; Bestimmung des
Permanganat-Index (ISO 8467:1993)

This European Standard was approved by CEN on 1994-11-03 and is identical to the ISO Standard as referred to.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration.

Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the Central Secretariat or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the Central Secretariat has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Denmark, Finland, France, Germany, Greece, Iceland, Ireland, Italy, Luxembourg, Netherlands, Norway, Portugal, Spain, Sweden, Switzerland and United Kingdom.

CEN

European Committee for Standardization
Comité Européen de Normalisation
Europäisches Komitee für Normung

Central Secretariat: rue de Stassart 36, B-1050 Brussels

Foreword

International Standard

ISO 8467:1993 Water quality; Determination of permanganate index,

which was prepared by ISO/TC 147 'Water quality' of the International Organization for Standardization, has been adopted by Technical Committee CEN/TC 230 'Water analysis' as a European Standard.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, and conflicting national standards withdrawn, by September 1995 at the latest.

In accordance with the CEN/GENELEC Internal Regulations, the following countries are bound to implement this European Standard:

Austria, Belgium, Denmark, Finland, France, Germany, Greece, Iceland, Ireland, Italy, Luxembourg, Netherlands, Norway, Portugal, Spain, Sweden, Switzerland and United Kingdom.

Endorsement notice

The text of the International Standard ISO 8467:1993 was approved by CEN as a European Standard without any modification.

Introduction

The permanganate index is a conventional measure of the contamination by organic and oxidizable inorganic matter in a water sample. It is primarily intended for judging the quality of potable water and raw waters, such as well and surface waters. More heavily contaminated waters may be analysed, provided an appropriate pre-dilution step is adopted. The permanganate index can be determined for waters containing less than 500 mg/l of chloride ion. Reducing compounds such as iron(II) salts, nitrites or hydrogen sulfide may contribute to the permanganate index but are not classified as impurities.

The permanganate index cannot be considered as a measure of the theoretical oxygen demand or the total content of organic matter. Many organic compounds are only partially oxidized in this test, as oxidation is generally incomplete. Volatile matter that evaporates before the addition of permanganate is not included.

The method is not recommended for determining organic load in waste waters; for this purpose the chemical oxygen demand should be determined as described in ISO 6060:1989, *Water quality — Determination of the chemical oxygen demand*.

The method is simple and convenient for surveying the quality of a large number of water samples.

1 Scope

This International Standard specifies a method for the determination of the permanganate index of water. It is primarily intended for water for human consumption and domestic use, drinking water, natural mineral water, well and table water, as well as water from swimming pools. It is used for the determination of the parameter "oxidizability". It is applicable to waters having a chloride ion concentration of less than 300 mg/l. Samples having a permanganate index over 10 mg/l should be diluted before analysis. The lower limit of the optimum range of the test is 0,5 mg/l.

2 Normative reference

The following standard contains provisions which, through reference in this text, constitute provisions of this International Standard. At the time of publication, the edition indicated was valid. All standards are subject to revision, and parties to agreements based on this International Standard are encouraged to investigate the possibility of applying the most recent edition of the standard indicated below. Members of IEC and ISO maintain registers of currently valid International Standards.

ISO 385-1:1984, *Laboratory glassware — Burettes — Part 1: General requirements.*

3 Definition

3.1 permanganate index (of water): The mass concentration of oxygen equivalent to the amount of permanganate ion consumed when a water sample is treated with that oxidant under defined conditions.

4 Principle

Heating of a sample in a boiling water-bath with a known amount of potassium permanganate and sulfuric acid for a fixed period of time (10 min). Re-

duction of part of the permanganate by oxidizable material in the sample and determination of the consumed permanganate by addition of an excess of oxalate solution, followed by titration with permanganate.

NOTE 1 The suggested maximum permanganate index of 10 mg/l is equivalent to a consumption of approximately 60 % of the added permanganate by the non-diluted sample.

5 Reagents

During the analysis, use only reagents of recognized analytical grade and only distilled water, water from a reverse osmosis or water of equivalent purity. Do not use deionized water from an organic ion exchanger.

NOTE 2 Non-reducing water can be prepared as follows. Add 10 ml of sulfuric acid (5.3) and a small excess of the potassium permanganate stock solution (5.6) to 1 litre of distilled water. Distil in an all-glass apparatus and discard the first 100 ml of distillate. Store in a glass bottle with a glass stopper.

The volume of the consumed potassium permanganate standard solution V_0 (see 8.4) shall not exceed 0,1 ml, otherwise the procedure shall be repeated or a water of lower organic content shall be used.

5.1 Sulfuric acid, $\rho(\text{H}_2\text{SO}_4) = 1,84 \text{ g/ml}$, 18 mol/l.

5.2 Sulfuric acid, $c(\text{H}_2\text{SO}_4) = 7,5 \text{ mol/l}$.

Add slowly, while stirring continuously, 420 ml of sulfuric acid (5.1) to 500 ml of water. Allow to cool and dilute to 1 litre.

5.3 Sulfuric acid, $c(\text{H}_2\text{SO}_4) = 2 \text{ mol/l}$.

Add slowly, while stirring continuously, 110 ml of sulfuric acid (5.1) to about 500 ml of water. Slowly add potassium permanganate solution (5.7) until a faint pink colour persists. Allow to cool, dilute with water to 1 litre and mix.