

**ÜHEKORDSELT KASUTATAVAD MEDITSIINILISED
KINDAD. OSA 3: BIOLOOGILISE HINDAMISE NÕUDED JA
KATSETAMINE**

**Medical gloves for single use - Part 3: Requirements
and testing for biological evaluation**

EESTI STANDARDI EESSÕNA**NATIONAL FOREWORD**

See Eesti standard EVS-EN 455-3:2015 sisaldab Euroopa standardi EN 455-3:2015 ingliskeelset teksti.	This Estonian standard EVS-EN 455-3:2015 consists of the English text of the European standard EN 455-3:2015.
Standard on jõustunud sellekohase teate avaldamisega EVS Teatajas.	This standard has been endorsed with a notification published in the official bulletin of the Estonian Centre for Standardisation.
Euroopa standardimisorganisatsioonid on teinud Euroopa standardi rahvuslikele liikmetele kättesaadavaks 01.04.2015.	Date of Availability of the European standard is 01.04.2015.
Standard on kättesaadav Eesti Standardikeskusest.	The standard is available from the Estonian Centre for Standardisation.

Tagasisidet standardi sisu kohta on võimalik edastada, kasutades EVS-i veebilehel asuvat tagasiside vormi või saates e-kirja meiliaadressile standardiosakond@evs.ee.

ICS 11.140

Standardite reprodutseerimise ja levitamise õigus kuulub Eesti Standardikeskusele

Andmete paljundamine, taastekitamine, kopeerimine, salvestamine elektroonsesse süsteemi või edastamine ükskõik millises vormis või millisel teel ilma Eesti Standardikeskuse kirjaliku loata on keelatud.

Kui Teil on küsimusi standardite autorikaitse kohta, võtke palun ühendust Eesti Standardikeskusega:
Aru 10, 10317 Tallinn, Eesti; koduleht www.evs.ee; telefon 605 5050; e-post info@evs.ee

The right to reproduce and distribute standards belongs to the Estonian Centre for Standardisation

No part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying, without a written permission from the Estonian Centre for Standardisation.

If you have any questions about copyright, please contact Estonian Centre for Standardisation:

Aru 10, 10317 Tallinn, Estonia; homepage www.evs.ee; phone +372 605 5050; e-mail info@evs.ee

English Version

Medical gloves for single use - Part 3: Requirements and testing for biological evaluation

Gants médicaux non réutilisables - Partie 3 : Exigences et
essais pour évaluation biologique

Medizinische Handschuhe zum einmaligen Gebrauch - Teil
3: Anforderungen und Prüfung für die biologische
Bewertung

This European Standard was approved by CEN on 24 January 2015.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION
COMITÉ EUROPÉEN DE NORMALISATION
EUROPÄISCHES KOMITEE FÜR NORMUNG

CEN-CENELEC Management Centre: Avenue Marnix 17, B-1000 Brussels

Contents	Page
Foreword.....	3
Introduction.....	5
1 Scope.....	6
2 Normative references.....	6
3 Terms and definitions.....	6
4 Requirements.....	7
4.1 General.....	7
4.2 Chemicals.....	7
4.3 Endotoxins.....	8
4.4 Powder-free gloves.....	8
4.5 Proteins, leachable.....	8
4.6 Labelling.....	8
5 Test methods.....	9
5.1 Endotoxins.....	9
5.2 Powder.....	9
5.3 Proteins, leachable.....	9
6 Test report.....	9
Annex A (normative) Method for the determination of aqueous extractable proteins in natural rubber gloves using the modified Lowry assay.....	10
Annex B (informative) Immunological methods for the measurements of natural rubber latex allergens.....	20
Annex C (informative) Amino acid analysis (AAA) by high pressure liquid chromatography (HPLC).....	26
Annex ZA (informative) Relationship between this European Standard and the Essential Requirements of EU Directive 93/42/EEC Medical Devices.....	34

Foreword

This document (EN 455-3:2015) has been prepared by Technical Committee CEN/TC 205 "Non-active medical devices", the secretariat of which is held by DIN.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by October 2015 and conflicting national standards shall be withdrawn at the latest by October 2015.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN [and/or CENELEC] shall not be held responsible for identifying any or all such patent rights.

This document supersedes EN 455-3:2006.

This document has been prepared under a mandate given to CEN by the European Commission and the European Free Trade Association, and supports essential requirements of EU Directive(s).

For relationship with EU Directive(s), see informative Annex ZA which is an integral part of this document.

With respect to EN 455-3:2006 the following changes are:

- a) standard was specified to relevant parts of EN ISO 10993 for Biological evaluation of medical devices;
- b) normative references revised;
- c) EN 980 was replaced by EN ISO 15223-1;
- d) subclause 4.2 "chemicals" was specified;
- e) subclause 4.4 specified as "powder-free gloves";
- f) level "As Low As Reasonably Practicable" (ALARP) removed in the whole standard;
- g) subclause 4.6 "labelling" specified;
- h) symbol for products containing natural latex (Figure 1) removed;
- i) the references in Annex B revised;
- j) Correspondence between this European Standard and Directive 89/686/EEC on Personal Protective Equipment made (see Annex ZA).

EN 455 consists of the following parts under the general title "*Medical gloves for single use*":

- *Part 1: Requirements and testing for freedom from holes;*
- *Part 2: Requirements and testing for physical properties;*
- *Part 3: Requirements and testing for biological evaluation;*
- *Part 4: Requirements and testing for shelf life determination.*

According to the CEN-CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece,

Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and the United Kingdom.

This document is a preview generated by EVS

Introduction

Adverse reactions to proteins in latex products have been reported over several years in variable rates of prevalence. Additionally, adverse reactions due to chemicals, lubricants, sterilization residues, pyrogens or other residues are described in the scientific literature. Adverse reactions are most often reported due to gloves made from natural rubber latex, but some of the reactions can also be seen due to gloves made from synthetic polymers.

EN ISO 10993 specifies requirements and test methods for biological evaluation of medical devices. However it does not specifically address adverse reactions that can result from the use of medical gloves (e.g, immediate type allergies). These adverse reactions occur to specific allergens that can be present in gloves. Several factors contribute to the risk of reaction:

- a) the duration and frequency of skin contact with gloves;
- b) the exposure to the allergens through direct contact to mucosa and skin (especially when not intact) and by inhalation of particles;
- c) the occlusive nature of the glove/skin interaction during glove use.

This part of EN 455 gives requirements and test methods for evaluation of the biological safety of medical gloves as part of a risk management process, in accordance with EN ISO 10993.

1 Scope

This part of EN 455 specifies requirements for the evaluation of biological safety for medical gloves for single use. It gives requirements for labelling and the disclosure of information relevant to the test methods used.

2 Normative references

The following documents, in whole or in part, are normatively referenced in this document and are indispensable for its application. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

EN 1041:2008+A1:2013, *Information supplied by the manufacturer of medical devices*

EN ISO 10993-1:2009, *Biological evaluation of medical devices - Part 1: Evaluation and testing within a risk management process (ISO 10993-1:2009)*

EN ISO 10993-5:2009, *Biological evaluation of medical devices - Part 5: Tests for in vitro cytotoxicity (ISO 10993-5:2009)*

EN ISO 10993-10:2013, *Biological evaluation of medical devices -Part 10: Tests for irritation and skin sensitization (ISO 10993-10:2010)*

EN ISO 14971:2012, *Medical devices - Application of risk management to medical devices (ISO 14971:2007, Corrected version 2007-10-01)*

EN ISO 15223-1:2012, *Medical devices - Symbols to be used with medical device labels, labelling and information to be supplied - Part 1: General requirements (ISO 15223-1:2012)*

EN ISO 21171:2006, *Medical gloves - Determination of removable surface powder (ISO 21171:2006)*

European Pharmacopoeia, General chapter 2.6.14 Bacterial Endotoxins: publisher EDQM - Council of Europe; 7 allée Kastner, CS 30026, F-67081 Strasbourg; France <http://www.edqm.eu/>

3 Terms and definitions

For the purposes of this document, the following terms and definitions apply.

3.1 chemicals
substances added or formed during any step of the manufacturing process or in storage which may be available in the final product

Note 1 to entry: These can include lubricants, chemical coatings and sterilizing agents. Several chemical ingredients are commonly used during processing of gloves, some of them are known to cause type IV allergic reactions. The type and amount of residual chemicals added and finally present are variable.

3.2 endotoxins
lipo-polysaccharides originating from the outer cell-membrane of Gram-negative bacteria

Note 1 to entry: Endotoxins are one type of pyrogen. Sources of endotoxins can include bacterial contamination of the raw materials, especially the process water used during manufacturing and manual handling of the gloves.

3.3 powder
all water insoluble material on the surface of a glove that is removed by washing under the conditions of the test